



# CALIBER ANALYTICAL SERVICES

## Certificate of Analysis

[Redacted]

Date Received: 10/13/09 12:30

Date Issued: 10/20/09 10:57

Project: Lead in Water

Site Location: [Redacted]

SDG Number: 09101302

	Result	Unit	LLQ	MCL	Method	Prepared	Analyzed	Init.
<b>Field Sample ID: 1</b>	<b>Matrix: Drinking Water</b>					<b>Date Sampled: 10/08/09 18:40</b>	<b>Lab ID: 09101302-01</b>	
<b>Total Lead</b>								
Lead	4.0	ug/L	1	15	EPA 200.8	10/15/09	10/15/09 12:14	MEL
<b>Field Sample ID: 2</b>	<b>Matrix: Drinking Water</b>					<b>Date Sampled: 10/08/09 18:43</b>	<b>Lab ID: 09101302-02</b>	
<b>Total Lead</b>								
Lead	ND	ug/L	1	15	EPA 200.8	10/15/09	10/15/09 12:18	MEL

Notes/Qualifiers:

LLQ- Lowest Level of Quantitation

MCL - Maximum Contaminant Level. The highest level of a contaminant that is allowed in drinking water. MCLs are associated with regulated analytes. They are enforceable standards.

ND - Not Detected at a concentration greater than or equal to the LLQ.

The above analyses performed by Maryland State Certified Water Quality Laboratory #320.

Approved by: Matt Cohen  
QC Chemist